## BOSTER BIOLOGICAL TECHNOLOGY Co.,Ltd.

3942 B Valley Ave, Pleasanton, CA, 94566

Phone: 888-466-3604 Fax: 925-215-2184 Email: boster@bosterbio.com Web: www.bosterbio.com

## **Certificate of Analysis**

Product Name: Polyclonal Anti-Zebrafish Abcb4 Antibody

Catalog Number: DZ01669-1

| Description   |   |
|---------------|---|
| Lot No.       | 20DZ213D15  |
| Size          | 200μl/vial  |
| Form          | Liquid  |
| Concentration | 500μg/ml  |
| lg type       | Rabbit IgG  |
| Purification  | Immunogen affinity purified.  |
| Contents      | Each vial contains 20mM PBS, 50% glycerol, 0.02% NaN <sub>3</sub> . |

| Data of Indirect ELISA                       |          |          |           |            |  |  |
|--|----------|----------|-----------|------------|--|--|
| Coating concentration of antigen             | 0.1µg/ml | 0.1µg/ml | 0.1µg/ml  | 0.1µg/ml   |  |  |
| Working concentration of anti-Abcb4 antibody | 1μg/ml   | 0.1µg/ml | 0.01µg/ml | 0.001µg/ml |  |  |
| Working concentration of HRP conjugated goat | 0.2µg/ml | 0.2µg/ml | 0.2µg/ml  | 0.2µg/ml   |  |  |
| anti-rabbit IgG (Catalog # BA1054)           |          |          |           |            |  |  |
| O.D. values(450nm)                           | 2.563    | 2.311    | 0.857     | 0.178      |  |  |
| Antibody Status                              | Pass     |          |           |            |  |  |

| Data Evaluation Criteria                                     |          |          |  |  |  |
|--|----------|----------|--|--|--|
| Coating concentration of antigen                             | 0.1µg/ml | 0.1µg/ml |  |  |  |
| Working concentration of the primary antibody                | 1µg/ml   | 1µg/ml   |  |  |  |
| Working concentration of HRP conjugated goat anti-rabbit IgG | 0.2μg/ml | 0.2μg/ml |  |  |  |
| O.D. values (450nm)  | ≥ 1.0    | <1.0     |  |  |  |
| Antibody's affinity for the antigen                          | Strong   | Weak     |  |  |  |
| Antibody Status  | Pass     | Fail     |  |  |  |

## Storage

At -20°C for one year, at 4°C for one month. Avoid repeated freezing and thawing.

1

## Approved by:

开杨莉

Ms. Yangjing Yue Chief technician Quality Control System of ELISA

Date: April 20, 2020