



# PicoKine™ ELISA

**Catalog number: EK7107**

For the quantitation of **Human HSPA8** concentrations in Cell lysates, Serum, Tissue

This package insert must be read in its entirety before using this product. For research use only. Not for use in diagnostic procedures.

## HSC70 ELISA Kit

Catalog Number: EK7107

### Introduction

Boster's ELISA Kit is for the detection of human Hsc70 in cell lysates, tissue extracts, and serum samples. Each kit contains sufficient components to quantitate the Hsc70 concentration in up to 40 samples, tested in duplicate.

### Overview

<b>Product Name</b>	HSC70 ELISA Kit
<b>Reactive Species</b>	Human
<b>Size</b>	96wells/kit, with removable strips.
<b>Description</b>	Colorimetric detection of HSC70. 96wells/kit, with removable strips.
<b>Sensitivity</b>	1.54 ng/ml *The sensitivity or the minimum detectable dose (MDD) is the lower limit of target protein that can be detected by the kit. It is determined by adding two standard deviations to the mean O.D. value of twenty (20) blank wells and calculating the corresponding concentration.
<b>Detection Range</b>	2.34 - 150 ng/mL
<b>Storage Instructions</b>	Store at 4°C.
<b>Uniprot ID</b>	P11142
<b>Specificity</b>	Natural and recombinant Human HSPA8
<b>Cross Reactivity</b>	There is no detectable cross-reactivity.

### Kit Components/Materials Provided

Description	Quantity
Anti-Hsc70 Immunoassay Plate	1 Plate
5X Hsc70 Extraction Reagent	1 vial/10 ml
Recombinant Hsc70 Standard	2 vials
Standard and Sample Diluent	1 vial/ 50 ml
10X Wash Buffer Concentrate	1 vial/100 ml
Anti-Hsc70 Biotinylated Antibody Concentrate	1 vial/150 µl

Phone: 888-466-3604

Fax: 925-215-2184

Email: [support@bosterbio.com](mailto:support@bosterbio.com)

Web: [www.bosterbio.com](http://www.bosterbio.com)

Anti-Hsc70 Biotinylated Antibody Diluent	1 vial/ 13 ml
Streptavidin: HRP Concentrate	1 vial/150 $\mu$ l
Streptavidin: HRP Diluent	1 vial/ 13 ml
TMB Substrate	1 vial/ 13 ml
Stop Solution	1 vial/ 13 ml

## Required Materials That Are Not Supplied

---

1. Ultra pure water.
2. Additional reagents and materials for cell lysate and tissue extract preparation, including protease inhibitors.
3. Precision pipettors, with disposable plastic tips.
4. Polypropylene or polyethylene tubes to prepare samples – do not use polystyrene, polycarbonate or glass tubes.
5. A container to prepare 1X Wash Buffer.
6. A wash bottle or an automated 96-well plate washer.
7. Disposable reagent reservoirs.
8. A standard microtiter plate reader for measuring absorbance at 450 nm.
9. Adhesive plate sealers.

## Assay Overview

---

1. Prepare Standard and samples in Standard and Sample Diluent.
2. Add 100  $\mu$ L of Standard or sample to appropriate wells.
3. Cover plate with Plate Sealer and incubate at room temperature (20-25°C) for 1 hour.
4. Wash plate four times with 1X Wash Buffer.
5. Add 100  $\mu$ L of Biotinylated Antibody Working Solution to each well.
6. Cover plate with Plate Sealer and incubate at room temperature for 1 hour.

7. Wash plate four times with 1X Wash Buffer.
8. Add 100  $\mu$ L of Streptavidin-HRP Working Solution to each well.
9. Cover plate with Plate Sealer and incubate at room temperature for 30 minutes.
10. Wash plate four times with 1X Wash Buffer.
11. Add 100  $\mu$ L of TMB Substrate to each well.
12. Develop the plate in the dark at room temperature for 30 minutes.
13. Stop reaction by adding 100  $\mu$ L of Stop Solution to each well.
14. Measure absorbance on a plate reader at 450 nm.

## Submit a product review to Biocompare.com

Submit a review of this product to [Biocompare.com](http://Biocompare.com) to receive a \$20 Amazon giftcard! Your reviews help your fellow scientists make the right decisions. Thank you for your contribution.

